## Certificate of Analysis



## Cathode Buffer Container 3500 Series for Protein Quality Analysis

Product No.

A31279

Lot No.

2007019

Date of Manufacture

27JUL2020

**Expiration Date** 

TEST

02APR2021

PHYSICAL INSPECTION

Material is clear, colorless and free of particulates

Appearance - Conformance

**Pass** 

RESULT

PH TEST

Measured at 25.0°C ± 0.3°C

pH Value

7.6 - 8.2

7.930

**CONDUCTIVITY TEST** 

Measured at 25.0°C ± 0.1°C

Conductivity

1500 - 2000 µS/cm

**SPECIFICATION** 

1765 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Life Technologies 5781 Van Allen Way Carlsbad, CA, USA 92008 www.thermofisher.com For inquiries, contact us at cofarequests@thermofisher.com Issued 28JUL20