

Certificate of Analysis

ThermoFisher
SCIENTIFIC


Cathode Buffer Container 3500 Series for Protein Quality Analysis

Product No. **A31279**
Lot No. **2102020**
Date of Manufacture **15FEB2021**
Expiration Date **06OCT2021**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.959
CONDUCTIVITY TEST		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1773 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 17FEB2021