


POROS ETHYL 50mL

Product No. **A32556**
 Lot No. **HET-012**
 Date of Manufacture **08AUG2018**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Slurry		Pass
DYNAMIC BINDING CAPACITY SOP 100043804		
5% Breakthrough with BSA	17.0 - 34.0 mg/mL	21.24 mg/mL
PROTEIN STANDARD SEPARATION SOP 100043803		
Lysozyme Conductivity	155.0 - 175.0 mS/cm	167.7930 mS/cm
Chymotrypsinogen A Conductivity	96.0 - 122.0 mS/cm	118.5680 mS/cm
FTIR ANALYSIS SOP BED3019		

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system.
 Site: Bedford, MA, USA**



Life Technologies
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at
cofarequests@thermofisher.com

**Quality Assurance
 Issued 08MAR19**


POROS ETHYL 50mL

Product No. **A32556**
 Lot No. **HET-012**
 Date of Manufacture **08AUG2018**

TEST	SPECIFICATION	RESULT
Compares to Reference Spectrum		Pass
ENDOTOXIN TEST SOP BED3069		
Endotoxin Level	< = 1.0 EU/mL	< 0.500 EU/mL
BIOBURDEN TEST SOP 4371392		
Total Bioburden	< = 100 CFU/mL	< 2.32 CFU/mL

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA



Life Technologies
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 08MAR19