

Certificate of Analysis

thermo**scientific**

POROS BENZYL 1,000mL

Product No. **A32560**
Lot No. **HBZ-012**
Date of Manufacture **13JUN2020**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Slurry		Pass
DYNAMIC BINDING CAPACITY SOP 100043804		
5% Breakthrough with Lysozyme	16.0 - 32.0 mg/mL	29.54 mg/mL
PROTEIN STANDARD SEPARATION SOP 100043803		
Ribonuclease A Conductivity	148.0 - 178.0 mS/cm	156.3080 mS/cm
Lysozyme Conductivity	108.0 - 136.0 mS/cm	120.8470 mS/cm
FTIR ANALYSIS SOP BED3019		

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA**

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Sandra Gil
Quality Assurance
Issued 20JUL20

Certificate of Analysis

thermo scientific

POROS BENZYL 1,000mL

Product No. A32560
Lot No. HBZ-012
Date of Manufacture 13JUN2020

TEST	SPECIFICATION	RESULT
Compares to Reference Spectrum		Pass
ENDOTOXIN TEST SOP BED3069		
Endotoxin Level	≤ 1.0 EU/mL	< 0.500 EU/mL
BIOBURDEN TEST SOP 4371392		
Total Bioburden	≤ 100 CFU/mL	< 2.64 CFU/mL

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

*Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA*

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Sandra Gil
Quality Assurance
Issued 20JUL20