

Certificate of Analysis

thermoscientific

POROS BENZYL 1,000mL

Product No. **A32560**
Lot No. **HBZ-015**
Date of Manufacture **20NOV2020**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Slurry		Pass
DYNAMIC BINDING CAPACITY SOP 100043804		
5% Breakthrough with Lysozyme	16.0 - 32.0 mg/mL	25.72 mg/mL
PROTEIN STANDARD SEPARATION SOP 100043803		
Ribonuclease A Conductivity	148.0 - 178.0 mS/cm	160.8260 mS/cm
Lysozyme Conductivity	108.0 - 136.0 mS/cm	123.4710 mS/cm
FTIR ANALYSIS SOP BED3019		

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


**Quality Assurance
Issued 11JAN2021**

Certificate of Analysis

thermo scientific

POROS BENZYL 1,000mL

Product No. A32560
Lot No. HBZ-015
Date of Manufacture 20NOV2020

TEST	SPECIFICATION	RESULT
Compares to Reference Spectrum		Pass
ENDOTOXIN TEST SOP BED3069		
Endotoxin Level	<= 1.0 EU/mL	<0.551 EU/mL
BIOBURDEN TEST SOP 4371392		
Total Bioburden	<= 100 CFU/mL	<2.64 CFU/mL

Pharmaceutical Grade Reagent. For Manufacturing and Laboratory Use only.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: 2
Preston Court Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 11JAN2021