

POROS BENZYL ULTRA 50mL

Product No. **A32568**
 Lot No. **HBZU-153**
 Date of Manufacture **25APR2017**

| TEST | SPECIFICATION | RESULT |
|---|---------------------|----------------|
| PHYSICAL INSPECTION | | |
| Labels Correct and Intact | | Pass |
| Caps Sealed | | Pass |
| PRODUCT APPEARANCE | | |
| White to Off-White Slurry | | Pass |
| DYNAMIC BINDING CAPACITY SOP 100043804 | | |
| 5% Breakthrough with Lysozyme | 25.0 - 41.0 mg/mL | 34.88 mg/mL |
| PROTEIN STANDARD SEPARATION SOP 100043803 | | |
| Ribonuclease A Conductivity | 125.0 - 153.0 mS/cm | 138.2260 mS/cm |
| Lysozyme Conductivity | 65.0 - 100.0 mS/cm | 86.0010 mS/cm |
| FTIR ANALYSIS SOP BED3019 | | |

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA



Life Technologies
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 14SEP17

POROS BENZYL ULTRA 50mL

Product No. **A32568**
 Lot No. **HBZU-153**
 Date of Manufacture **25APR2017**

| TEST | SPECIFICATION | RESULT |
|--------------------------------------|----------------|---------------|
| Compares to Reference Spectrum | | Pass |
| ENDOTOXIN TEST SOP BED3069 | | |
| Endotoxin Level | < = 1.0 EU/mL | < 0.500 EU/mL |
| BIOBURDEN TEST SOP 4371392 | | |
| Total Bioburden | < = 100 CFU/mL | 3.30 CFU/mL |

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA



Life Technologies
 5781 Van Allen Way
 Carlsbad, CA, USA 92008
www.thermofisher.com
 For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 14SEP17