Certificate of Analysis



Cathode Buffer Container CE-SeqStudio

Product No.

A33401

Lot No.

2104067

Date of Manufacture

04MAY2021

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
PH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.991
CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1802 μS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Quality Assurance Issued 05MAY2021