Certificate of Analysis

thermoscientific

POROS CAPRYLATE MMCEX 50 ML

Product No. A51049

Lot No. MMCEX-002

Date of Manufacture **08AUG2023**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
Labels Correct and Intact		Pass
Caps Sealed		Pass
PRODUCT APPEARANCE		
White to Off-White Slurry		Pass
FRONTAL CAPACITY SOP 100092692#		
50% Breakthrough with lysozyme	>= 20.0 mg/mL	27.590 mg/mL
EXCHANGE CAPACITY SOP 100092035		
Exchange Capacity	45.0 - 65.0 μmo/mL	56.320 μmo/mL

For Research Use Only.

Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Quality Assurance Issued 24JUN2024

Certificate of Analysis

thermoscientific

POROS CAPRYLATE MMCEX 50 ML

Product No. A51049

Lot No. MMCEX-002

Date of Manufacture **08AUG2023**

TEST	SPECIFICATION	RESULT
FTIR ANALYSIS		
SOP BED3019		

Compare to Reference Spectrum Pass

ENDOTOXIN TEST SOP BED3069

JOI DED3009

Endotoxin Level <= 1.0 EU/mL < 0.622 EU/mL

BIOBURDEN TEST

SOP 4371392

Total Bioburden <= 100 CFU/mL 10 CFU/mL

For Research Use Only.

Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Quality Assurance Issued 24JUN2024