

# Certificate of Analysis

## ViralSEQ™ dPCR Lentivirus Physical Titer Kit

**Product No.** A59317  
**Lot No.** 2311001  
**Date of Manufacture** 12DEC2023  
**Expiration Date** 05OCT2024

### KIT CONFIGURATIONS:

COMPONENT DESCRIPTION	LOT NO.
100122791 Tube, Absolute Q RT-dPCR Master Mix	2836361
100122788 Tube, dPCR Lentivirus Physical Titer Assay Mix	2311001
100102063 Tube, Nuclease Free Water	2311012

TEST	SPECIFICATIONS	RESULT
------	----------------	--------

QC Method: The positive results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

**Lentiviral RNA (FAM)** Pass Functional Test Pass

**IPC (VIC)** 90 – 220 cp/μL Pass

QC Method: The negative results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

**No Template Control (NTC)** Pass Functional Test Pass

*For Research Use Only. Not for use in diagnostic procedures.*

Manufactured in compliance with our ISO 13485 certified quality management system.

**Site:** Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



Quality Assurance  
Issued 12APR2024