

# Certificate of Analysis



## RNase-free Tips, 200 µl size (8 x 12 racks)

Product No. AM12650  
Lot No. 19220402  
Date of Manufacture 10JUL2020

TEST	SPECIFICATION	RESULT
<b>Non-Pyrogenic Testing</b>		
Met the criteria established in the current version of ANSI/AAMI ST 72.		Pass
<b>Sterilization</b>		
Radiation		
Sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137.	$\geq 10^{-6}$ Sterility Assurance Level	Pass
<b>Human DNA Free</b>		
Tested by PCR method		
Human DNA	$\leq 5\text{pg}$	Pass
<b>DNase/RNase Free</b>		
Tested by nuclease assay method		
DNase	$\leq 10^{-5}$ Kunitz units/uL	Pass
RNase	$\leq 10^{-9}$ Kunitz units/uL	Pass
<b>Quality Control Testing</b>		
Visual Inspection		Pass
Functional Test		Pass
Dimensional Test		Pass
Accuracy Test		Pass
<b><i>For Research Use Only. Not for use in diagnostic procedures.</i></b>		

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. **Site:** Austin, TX, USA

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

Quality Assurance  
Issued 23JUN2021

FORM-00013G  
Doc: 100054706B