

Certificate of Analysis



RNase-free Tips, 200 µl size (8 x 12 racks)

Product No. AM12650
Lot No. 19720402
Date of Manufacture 15JUL2020

TEST	SPECIFICATION	RESULT
Non-Pyrogenic Testing		
Met the criteria established in the current version of ANSI/AAMI ST 72.		Pass
Sterilization		
Radiation		
Sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137.	$\geq 10^{-6}$ Sterility Assurance Level	Pass
Human DNA Free		
Tested by PCR method		
Human DNA	$\leq 5\text{pg}$	Pass
DNase/RNase Free		
Tested by nuclease assay method		
DNase	$\leq 10^{-5}$ Kunitz units/uL	Pass
RNase	$\leq 10^{-9}$ Kunitz units/uL	Pass
Quality Control Testing		
Visual Inspection		Pass
Functional Test		Pass
Dimensional Test		Pass
Accuracy Test		Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. **Site:** Austin, TX, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 21JUN2021

FORM-00013G
Doc: 100054706B