

Certificate of Analysis



RNase-free Tips, 200 µl size (8 x 12 racks)

Product No. AM12650
Lot No. 22280602
Date of Manufacture 14JUL2022

| TEST | SPECIFICATION | RESULT |
|--|-----------------------------------|--------|
| Endotoxin (Pyrogen) | | |
| Endotoxin (Pyrogen) | <.06 EU/mL | Pass |
| PCR Inhibitor | | |
| PCR Inhibitors | Not present | Pass |
| Human and mouse DNA | | |
| Tested by PCR method | | |
| Human and mouse DNA | ≤30pg | Pass |
| DNase/RNase Free | | |
| Tested by nuclease assay method | | |
| DNase | ≤10 ⁻⁷ Kunitz units/uL | Pass |
| RNase | ≤10 ⁻⁹ Kunitz units/uL | Pass |
| Adenosine-5'-triphosphate (ATP) | | |
| ATP | < 10 ⁻¹³ mg/µl | Pass |

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. **Site:** Austin, TX, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 13SEP2022

FORM-00013G
Doc: 100054706C