

Certificate of Analysis



RNase-free Tips, 200 µl size (8 x 12 racks)

Product No. AM12650
Lot No. 32320402
Date of Manufacture 23NOV2020

TEST	SPECIFICATION	RESULT
Non-Pyrogenic Testing		
Met the criteria established in the current version of ANSI/AAMI ST 72.		Pass
Sterilization		
Radiation		
Sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137.	$\geq 10^{-6}$ Sterility Assurance Level	Pass
Human DNA Free		
Tested by PCR method		
Human DNA	$\leq 5\text{pg}$	Pass
DNase/RNase Free		
Tested by nuclease assay method		
DNase	$\leq 10^{-5}$ Kunitz units/uL	Pass
RNase	$\leq 10^{-9}$ Kunitz units/uL	Pass
Quality Control Testing		
Visual Inspection		Pass
Functional Test		Pass
Dimensional Test		Pass
Accuracy Test		Pass
<i>For Research Use Only. Not for use in diagnostic procedures.</i>		

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. **Site:** Austin, TX, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 24AUG2021

FORM-00013G
Doc: 100054706B