

## Certificate of Analysis

### Quick Apoptotic DNA Ladder Detection Kit

Product No. KHO1021  
Lot No. 76089587  
Expiration Date 04Aug2020

---

#### Specifications

- Visual: DNA Suspension Buffer Clear Orange Solution
- Visual: Enzyme B appearance normal after lyophilization
- Assay: Time study with camptothecin induced Jurkat cells

#### Results

- Visual: Both confirmed
- Assay: Passed DNA Ladder observed after 6 and 12 hr. induction. No ladder without induction

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



---

Kelly Mitchell-Haloskey  
Manager, Quality  
Issued on 16 Aug 2019