

Certificate of Analysis

Quick Apoptotic DNA Ladder Detection Kit

Product No. KHO1021
Lot No. 76089587A
Expiration Date 04Aug2020

Specifications

- Visual: DNA Suspension Buffer Clear Orange Solution
- Visual: Enzyme B appearance normal after lyophilization
- Assay: Time study with camptothecin induced Jurkat cells

Results

- Visual: Both confirmed
- Assay: Passed DNA Ladder observed after 6 and 12 hr. induction. No ladder without induction

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Kelly Mitchell-Haloskey
Manager, Quality
Issued on 13 Nov 2019