

PRODUCT CERTIFICATE

October 19, 2017

The following information represents Product Certification for:

Catalog Number: **170365N**
Lot Number: **GB08002**
Description: **2mL Serological Pipette**
Date of Manufacture: **02/2017**
Use By Date: **02/2022**

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to the relevant specifications:

| <u>Test Description</u> | <u>Requirements</u> | <u>Result</u> |
|-----------------------------------|--|----------------------|
| Accuracy | ± 2% at full volume | Complies |
| USP <85> Bacterial Endotoxin Test | < 0.5 EU/mL (<20 EU/Part) | Complies |
| RNase Test | ≤ 1.9 X 10 ⁻¹⁰ Kunitz unit/μL | Complies |
| DNase Test | ≤ 8 X 10 ⁻⁷ Kunitz unit/μL | Complies |
| Human DNA Test | < 0.6 ng hDNA | Complies |

Materials were tested and shown to be non-cytotoxic as part of our initial product validation per ISO 10993-5, MTT Cytotoxicity Test.

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation" with SAL 10⁻⁶.

Quality assurance of the product is performed in accordance with the requirements of Quality management System ISO 9001:2008 "Quality management system - Requirements".

Sincerely,



Keith Strassman
Raw Material and Compliance Specialist
Thermo Fisher Scientific
Labware and Specialty Plastics
Certificate ROC@thermofisher.com