

Certificate of Analysis

ThermoFisher
SCIENTIFIC

CDP-Star Ready-to-Use Substrate 0.25mM

Product No. T2145
Lot No. 2006083
Date of Manufacture 30JUN2020

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	16 - 37 RLU	16 RLU
ENZYME FUNCTIONAL TEST		
Signal/Noise - Low	26.5 - 55.5 S/N	46.0 S/N
Signal/Noise - Medium	251 - 548 S/N	448 S/N
Signal/Noise - High	15050 - 31369 S/N	30424 S/N
HPLC PURITY		
% Purity by HPLC	>= 75.00 %	96.13 %

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Sandra Grl
Quality Assurance
Issued 22JUL20