

# Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC

## CDP-Star Ready-To-Use Substrate 0.25mM

Product No. T2145  
Lot No. 2011084  
Date of Manufacture 30JUN2020

TEST	SPECIFICATION	RESULT
<b>BACKGROUND</b>		
Background Luminescence (RLU)	<50 RLU	0.0 RLU
<b>% ACTIVITY</b>		
Signal Test/Signal STD	40 - 86%	68.4 %
<b>HPLC PURITY</b>		
% Purity by HPLC (concentrate)	>= 90.00 %	97.12 %

*For Research Use Only. Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:  
Bedford, MA, USA**

 Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Sandra Gil*  
**Quality Assurance**  
**Issued 04DEC2020**