

Certificate of Analysis

ThermoFisher
SCIENTIFIC


CDP-Star Ready-To-Use Substrate 0.25mM


Product No. T2145
Lot No. 2201088
Date of Manufacture 09SEP2021

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 50	0.0
% ACTIVITY		
Signal Test/Signal STD	40 - 86 %	66.8 %
HPLC PURITY		
% Purity by HPLC (concentrate)	>= 90.00 %	99.56 %

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Kyle Cimon (Feb 4, 2022 10:37 EST)

**Quality Assurance
Issued 04FEB2022**