

Certificate of Analysis



CDP-Star Ready-To-Use Substrate 0.25mM

Product No. T2145
Lot No. 2407096
Date of Manufacture 05FEB2024

TEST	SPECIFICATION	RESULT
Background		
Background Luminescence (RLU)	<= 50	0.3
% Activity		
Signal Test/Signal STD	40 - 86 %	70.4 %
HPLC Purity		
% Purity by HPLC (concentrate)	>= 90.00 %	95.26%

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Janelle Him
Quality Assurance
Issued 22AUG2024