

Certificate of Analysis

ThermoFisher
SCIENTIFIC


CDP-Star Ready-To-Use Substrate 0.25mM (100mL)

Product No. T2146
Lot No. 2101107
Date of Manufacture 05AUG2020

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 50 RLU	0.0 RLU
% ACTIVITY		
Signal Test/Signal STD	40 - 86 %	69.3 %
HPLC PURITY		
% Purity by HPLC (concentrate)	>= 90.00 %	97.12 %

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 26JAN2021