

# Certificate of Analysis



## 10X Assay Buffer

Product No. T2187  
Lot No. 2403099  
Date of Manufacture 26JUL2023

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	2 - 11 RLU	4.05 RLU

*For Research Use Only. Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Amy Fantasia*  
**Quality Assurance  
Issued 20MAR2024**