

# Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC


## CSPD RTU W/ SAPPHIRE-II 100 ML

Product No. T2210  
Lot No. 2008086  
Date of Manufacture 12MAY2020

TEST	SPECIFICATION	RESULT
<b>BACKGROUND</b>		
Background Luminescence (RLU)	$\leq 100$ RLU	1.8 RLU
<b>% ACTIVITY</b>		
Signal Test/Signal STD	6 - 16 %	11 %

*For Research Use Only. Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:  
Bedford, MA, USA**

 Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
**Quality Assurance  
Issued 24SEP2020**