

Certificate of Analysis

ThermoFisher
SCIENTIFIC

CSPD RTU W/ SAPPHIRE-II 100 ML


Product No. T2210
Lot No. 2009087
Date of Manufacture 12MAY2020

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	≤ 100 RLU	1.8 RLU
% ACTIVITY		
Signal Test/Signal STD	6 - 16 %	11 %

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:
Bedford, MA, USA**

 Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 08OCT2020