

# Certificate of Analysis



## CSPD RTU W/ SAPPHIRE-II 100 ML

Product No. T2210  
Lot No. 2407099  
Date of Manufacture 13MAY2024

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 100 RLU	2.7 RLU
% ACTIVITY		
Signal Test/Signal STD	6 - 16 %	10.0 %

**For Research Use Only.**  
**Not for use in diagnostic procedures.**

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:**  
**Bedford, MA, USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Amy Fantasia*  
**Quality Assurance**  
**Issued 23OCT2024**