

Certificate of Analysis



CDP STAR READY TO USE WITH SAPPHIRE II

Product No. T2214
Lot No. 2403135
Date of Manufacture 05JAN2024

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	≤ 100 RLU	2.7 RLU
% ACTIVITY		
Signal Test/Signal STD	31 - 74 %	45.0 %

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Bedford, MA, USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Amy Fantasia
**Quality Assurance
Issued 20MAR2024**