

Certificate of Analysis



CDP STAR READY TO USE WITH SAPPHIRE II

Product No. T2214
Lot No. 2404136
Date of Manufacture 14JUL2023

TEST	SPECIFICATION	RESULT
BACKGROUND		
Background Luminescence (RLU)	<= 100 RLU	0.6 RLU
% ACTIVITY		
Signal Test/Signal STD	31 - 74 %	39.1 %

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Janelle Him
Quality Assurance
Issued 07MAY2024