

# Certificate of Analysis



## CDP STAR READY TO USE WITH SAPPHIRE II

Product No. T2214  
Lot No. 2407138  
Date of Manufacture 30APR2024

| TEST                          | SPECIFICATION | RESULT   |
|-------------------------------|---------------|----------|
| BACKGROUND                    |               |          |
| Background Luminescence (RLU) | <= 100 RLU    | 11.5 RLU |
| % ACTIVITY                    |               |          |
| Signal Test/Signal STD        | 31 - 74 %     | 50.3 %   |

*For Research Use Only.  
Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:  
Bedford, MA, USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Amy Fantasia*  
**Quality Assurance  
Issued 23OCT2024**