	Document Owner Department: QC	BT-SPEC-0632
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NEOMYCIN N30 CT0033B		

Product Details

Product Name	NEOMYCIN
Disc Code	N30
Disc Concentration	30µg
CT Code	CT0033B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	No

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Staphylococcus aureus</i>	ATCC® 25923	18 – 26
<i>Escherichia coli</i>	ATCC® 25922	17 – 23


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: VET01 Performance standards for antimicrobial disk and dilution susceptibility tests for bacteria isolated from animals

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0632
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NEOMYCIN N30 CT0033B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document/ References	Update to current format. Update of current performance standards	Change control	BT-CC-2825