	Document Owner Department: QC	BT-SPEC-0703
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NORFLOXACIN NOR10 CT0434B		

Product Details

Product Name	NORFLOXACIN
Disc Code	NOR10
Disc Concentration	10µg
CT Code	CT0434B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	28 – 35
<i>Staphylococcus aureus</i>	ATCC® 25923	17 – 28
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	22 – 29
<i>Streptococcus pneumoniae</i>	ATCC® 49619	15 – 21

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	28 – 35
<i>Staphylococcus aureus</i>	ATCC® 29213	18 – 24
<i>Streptococcus pneumoniae</i>	ATCC® 49619	18 – 24
<i>Enterococcus faecalis</i>	ATCC® 29212	16 – 22


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- EUCAST Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0703
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NORFLOXACIN NOR10 CT0434B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document/ EUCAST table/ References	Update to current format/ Addition of <i>Enterococcus faecalis</i> ATCC® 29212 testing / Update of current performance standards	Change control	BT-CC-2887