

| | | |
|---|-------------------------------|--------------|
| ThermoFisher SCIENTIFIC | Document Owner Department: QC | BT-SPEC-0772 |
| | | Page 1 of 2 |
| OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION | | |
| CEFTAROLINE CPT30 CT1941B | | |

Product Details

| | |
|---|-------------|
| Product Name | CEFTAROLINE |
| Disc Code | CPT30 |
| Disc Concentration | 30µg |
| CT Code | CT1941B |
| Shelf Life (years) | 1 |
| CLSI Recommended Concentration | Yes |
| EUCAST Recommended Concentration | No |

Microbiological Performance Criteria

| CLSI | | |
|---------------------------------|------------------|------------|
| Organism | ATCC/NCTC Number | Range (mm) |
| <i>Escherichia coli</i> | ATCC®25922 | 26 – 34 |
| <i>Staphylococcus aureus</i> | ATCC®25923 | 26 – 35 |
| <i>Haemophilus influenzae</i> | ATCC®49247 | 29 – 39 |
| <i>Streptococcus pneumoniae</i> | ATCC®49619 | 31 – 41 |


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

| | | |
|---|-------------------------------|--------------|
|  | Document Owner Department: QC | BT-SPEC-0772 |
| | | Page 2 of 2 |
| OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION | | |
| CEFTAROLINE CPT30 CT1941B | | |

Revision History

| Section / Step | Description of Change | Reason for Change | Reference |
|--------------------------------|---|-------------------|------------|
| Entire document/ References | Update to current format/ Update to current performance standards | Change control | BT-CC-2820 |