	Document Owner Department: QC	MBD-BT-SPEC-0825
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		Revision 2
TEMOCILLIN TEM30 CT1952B		

## Product Details

<b>Product Name</b>	TEMOCILLIN
<b>Disc Code</b>	TEM30
<b>Disc Concentration</b>	30µg
<b>CT Code</b>	CT1952B
<b>Shelf Life (years)</b>	1
<b>CLSI Recommended Concentration</b>	No
<b>EUCAST Recommended Concentration</b>	Yes

## Microbiological Performance Criteria

Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 35218	19 – 28
<i>Staphylococcus aureus</i>	ATCC® 25923	6 – 12

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	16 – 22

## References


Tested in accordance with the current CLSI methodology, the control ranges are internally derived based on historical data.

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- EUCAST: Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	MBD-BT-SPEC-0825
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		Revision 2
TEMOCILLIN TEM30 CT1952B		

**Revision History**

Section / Step	Description of Change	Reason for Change	Reference
EUCAST table	Creation of EUCAST table for recommended strain with new limits	Change control	MOC-2023-0986