



CERTIFICATE OF COMPLIANCE LRQA

hereby declares that the certification assessment has demonstrated that

SureTect[™] Salmonella Species PCR Assay

Supplied by: Manufactured by:

Thermo Fisher Scientific Thermo Fisher Scientific

Wade Road, Basingstoke Ratastie 2 Hampshire, RG24 8PW Vantaa 01620

United Kingdom Finland

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method:

ISO 6579 – 1:2017 Microbiology of food chain – Horizontal method for detection, enumeration and serotyping of *Salmonella* – Part 1: Detection of *Salmonella* spp.

Scope: a broad range of foods and environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 9.1.

Certificate no.: 2022LR111 First approval date: 24 March 2023 Expiry date: 23 March 2027

Expiry date. 25 March 2027

ISSUED BY: LRQA Nederland B.V.

Rotterdam, The Netherlands

Certificate no.: 2022LR111 15-05-2023 Page 1 of 2





Year modification	Description modification	Date approval MVTC
2023	Optional automated	12 December 2023 – 59 th
	procedure for lysis and	MVTC meeting
	PCR setup. Equivalent to	
	manual lysis and PCR	
	setup procedure.	
2024	Inclusion of an optional	12 March 2024 – 60 th
	supplement format,	MVTC Meeting
	PrecisBlue, for buffered	
	peptone water certified	
	enrichments.	

Certificate no.: 2022LR111 15-05-2023 Page 2 of 2