

**CERTIFICATE OF ANALYSIS**

PRODUCT MA0114D
OXACILLIN 256 M.I.C.E. (10 Pack)

LOT NUMBER 1889822

EXPIRY DATE 2017.07.31

DATE OF MANUFACTURE 2016.06.28

Delivery/Customer information

Date Printed

2016.07.05

Delivery No.

Customer

Customer Order number

Microbiological Performance**Mueller-Hinton Agar with the addition of 2% w/v sodium chloride****Incubated at 35-37°C in air for 16-18 hours**

	MIC (µg/mL)	Limits (µg/mL)
<i>Staphylococcus aureus</i> ATCC®29213	0.18	0.09 - 0.50

Mueller-Hinton Agar**Incubated at 35-37°C in air for 16-18 hours**


<i>Enterococcus faecalis</i> ATCC®29212	12.00	6.00 - 32.00
---	-------	--------------

The result is expressed as the mean result rounded to the nearest half or full MIC value. Where the mean MIC value falls between two MIC values the highest value is reported.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted.


.....
Carol MacBride
Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Product Performance Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096