



# CERTIFICATE OF ANALYSIS

**PRODUCT** CM0097B  
CZAPEK DOX AGAR (MODIFIED) 500g

**LOT NUMBER** 3663842

**EXPIRY DATE** 2028.05.31

**DATE OF MANUFACTURE** 2023.04.25

**Delivery/Customer information**

Date Printed

2023.05.26

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	White powder	White powder
Colour on reconstitution	Off white	Off white
pH (25°C)	6.6	6.6 - 7.0
Clarity	Clear	Clear
Microbiological Performance	Control cfu	Test cfu
		Recovery Test % Description
<b>Aerobic incubation at 25°C for 5 days</b>		
<i>Aspergillus brasiliensis</i> ATCC®9642	1E+04-1E+06	White/yellow mycelia, black spores
<i>Penicillium chrysogenum</i> ATCC®9179	1E+04-1E+06	White mycelia, green spores
<i>Candida albicans</i> ATCC®10231	1E+04-1E+06	Cream cols
<i>Bacillus cereus</i> ATCC®10876	1E+04-1E+06	Irregular, white cols
<i>Candida albicans</i> ATCC®10231	1E+04-1E+06	White cols & chlamydospores

Control Media: Tryptone Soya Agar or Sabouraud Dextrose Agar, where appropriate

A satisfactory result is represented by growth from an inoculum of 1E+04 to 1E+06 colony-forming units (cfu).  
Refer to product specification for full details.



Tested by the Quality Control Laboratory  
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096



## CERTIFICATE OF ANALYSIS

**PRODUCT** CM0097B  
**LOT NUMBER** CZAPEK DOX AGAR (MODIFIED) 500g  
3663842

---

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.05.10

.....  
Carissa Courtney  
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory  
OXOID LIMITED  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxoid@thermofisher.com www.oxoid.com