

CERTIFICATE OF ANALYSIS

PRODUCT CM0533B
S S AGAR (MODIFIED) 500g

LOT NUMBER 6233008

EXPIRY DATE 2030.02.28

DATE OF MANUFACTURE 2025.02.14

Delivery/Customer information

Date Printed

2025.03.20

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Pink/red powder	Pink/red powder
Colour on reconstitution	Pink/red	Pink/red
pH (25°C)	7.3	7.1 - 7.5
Clarity	Clear	Clear

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Aerobic incubation at 37°C for 18-24 hours

Mixed cultures challenged with 1E+03 to 1E+05 cfu *Escherichia coli* ATCC®8739

Inoculation using diminishing sweep technique

NOTE:Salmonella colonies may only be visible in the secondary or tertiary bed of the mixed culture streak.

<i>Salmonella enteritidis</i> ATCC®13076	8.1E+04	-	-	Straw cols with black centre
<i>Salmonella typhimurium</i> ATCC®14028	6.2E+04	-	-	Straw cols with black centre
<i>Salmonella virchow</i> NCTC5742	9.5E+04	-	-	Straw cols with black centre
<i>Shigella sonnei</i> ATCC®25931	7.2E+04	-	-	Irregular, translucent pink cols
<i>Shigella boydii</i> NCTC11462	3.9E+04	-	-	Irregular, translucent pink cols
<i>Shigella flexneri</i> ATCC®12022	5.4E+04	-	-	Translucent pink cols

Pure cultures

<i>Salmonella typhimurium</i> ATCC®14028	56	69	123	Straw cols with black centre
<i>Shigella sonnei</i> ATCC®25931	72	75	104	Irregular, translucent pink cols
<i>Shigella dysenteriae</i> NCTC9721	39	23	59	Translucent pink cols
<i>Pseudomonas aeruginosa</i> ATCC®27853	100	98	98	Straw cols
<i>Proteus mirabilis</i> ATCC®12453	99	84	85	Straw cols, grey centre & no swarm
<i>Proteus mirabilis</i> ATCC®29906	60	57	95	Straw cols, grey centre & no swarm
<i>Escherichia coli</i> ATCC®11775	66	52	79	Pink cols
<i>Escherichia coli</i> ATCC®25922	33	0	0	No growth
<i>Enterococcus faecalis</i> ATCC®29212	>1E+04	0	-	No growth

Control Medium: Tryptone Soya Agar

For mixed cultures, a satisfactory result is represented by differential diagnostic reactions of Salmonellae and Shigellae strains and *Escherichia coli*.

For pure cultures, a satisfactory result is represented by recovery of Salmonellae and Shigellae strains equal to or greater than 70% of the control medium from an inoculum of 10-100 cfu.

For *Shigella dysenteriae* NCTC9721, a satisfactory result is represented by recovery equal to or greater than 50% of the control medium

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT	CM0533B	
	S S AGAR (MODIFIED)	500g
LOT NUMBER	6233008	

from an inoculum of 10-100 cfu.

For other strains, a satisfactory result is represented by recovery equal to or less than 100% of the control medium from an inoculum of 10-100 cfu.

Negative strains are inhibited or shall produce a negative diagnostic reaction.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.03.13

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com