



CERTIFICATE OF ANALYSIS

PRODUCT CM0785B
M17 MEDIUM 500g

LOT NUMBER 3503080

EXPIRY DATE 2024.06.30

DATE OF MANUFACTURE 2022.06.15

Delivery/Customer information

Date Printed

2022.07.21

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results		Specification	
Appearance	Straw powder		Straw powder	
Colour on reconstitution	Straw 2-3		Straw 2-3	
pH (25°C)	7.0		6.7 - 7.1	
Clarity	Clear		Clear	
Microbiological Performance	Control cfu	Test cfu	Recovery %	Test Description
Aerobic incubation at 37°C for 48 hours <i>Streptococcus thermophilus</i> ATCC®14485	48	45	94	White cols
Aerobic incubation at 25°C for 48 hours <i>Lact. lactic ss cremoris</i> ATCC®19257	34	25	74	White cols
Pour plate technique <i>Streptococcus thermophilus</i> ATCC®14485	89	88	99	White cols
Pour plate technique <i>Lact. lactic ss cremoris</i> ATCC®19257	64	48	75	White cols

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu).

Refer to product specification for full details.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096

**CERTIFICATE OF ANALYSIS**

PRODUCT	CM0785B	
	M17 MEDIUM	500g
LOT NUMBER	3503080	

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2022.07.12

.....
Carissa Courtney
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com