

CERTIFICATE OF ANALYSIS

PRODUCT CM0888B
BILE AESCULIN AGAR 500g

LOT NUMBER 6180804

EXPIRY DATE 2029.10.31

DATE OF MANUFACTURE 2024.09.30

Delivery/Customer information

Date Printed

2024.11.06

Delivery No.

Customer

Customer Order Number

Physical Characteristics

Results

Specification

Appearance
Colour on reconstitution
pH (25°C)
Clarity

Straw powder
Straw/brown
7.0
Clear

Straw powder
Straw/brown
6.9 - 7.3
Clear

Microbiological Performance

Control
cfu

Test
cfu

Recovery Test
% Description

Aerobic incubation at 37°C for 18 to 24 hours

<i>Enterococcus faecalis</i> ATCC®19433	31	29	94	Brown cols, aesculin hydrolysis
<i>Enterococcus faecalis</i> ATCC®29212	>1E+04	-	-	Brown cols, aesculin hydrolysis
<i>Enterococcus faecium</i> ATCC®19434	>1E+04	-	-	Brown cols, aesculin hydrolysis
<i>Enterococcus durans</i> ATCC®19432	>1E+04	-	-	Brown cols, aesculin hydrolysis
<i>Streptococcus bovis</i> ATCC®27960	>1E+04	-	-	Brown cols, aesculin hydrolysis
<i>Enterobacter aerogenes</i> ATCC®13048	>1E+04	-	-	Brown cols, aesculin hydrolysis
<i>Streptococcus pyogenes</i> ATCC®19615	>1E+04	0	-	No growth

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium from an inoculum of 10-100 colony-forming units (cfu). For positive test strains inoculated using diminishing sweep technique, a satisfactory result is represented by growth in accordance with the specification from an inoculum of 1E+04 to 1E+06 cfu. Negative strains are inhibited from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2024.10.14

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com