

**CERTIFICATE OF ANALYSIS**

PRODUCT CM0957B
ANAEROBE BASAL BROTH 500g

LOT NUMBER 3409400

EXPIRY DATE 2024.12.31

DATE OF MANUFACTURE 2021.12.16

Delivery/Customer information

Date Printed

2022.01.22

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw/green	Straw/green
pH (25°C)	7.0	6.6 - 7.0
Clarity	Clear	Clear
Microbiological Performance	Initial Inoculum	Test Result
Anaerobic incubation at 37°C for 48 hours		
<i>Peptostreptococcus anaerobius</i> ATCC®27337	30	Turbid growth
<i>Bacteroides thetaiotaomicron</i> ATCC®29148	1E+04-1E+06	Turbid growth
<i>Prevotella loescheii</i> ATCC®15930	1E+04-1E+06	Turbid growth
<i>Clostridium tetani</i> ATCC®9441	1E+04-1E+06	Turbid growth
<i>Clostridium sphenoides</i> ATCC®19403	1E+04-1E+06	Turbid growth
<i>Clostridium perfringens</i> ATCC®13124	1E+04-1E+06	Turbid growth
<i>Fusobacterium nucleatum</i> ATCC®10953	1E+04-1E+06	Turbid growth

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu). For other positive test strains, a satisfactory result is represented by visible growth in accordance with the specification from an inoculum of 1E+04 to 1E+06 cfu. Refer to product specification for full details.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2022.01.13

Carissa Courtney
Director Quality, EMEA



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT

CM0957B

ANAEROBE BASAL BROTH 500g

LOT NUMBER

3409400

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096