

**CERTIFICATE OF ANALYSIS**

**PRODUCT** CM0957B  
ANAEROBE BASAL BROTH 500g

**LOT NUMBER** 3725312

**EXPIRY DATE** 2026.09.30

**DATE OF MANUFACTURE** 2023.08.31

**Delivery/Customer information**

Date Printed

2023.10.12

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw/green	Straw/green
pH (25°C)	6.8	6.6 - 7.0
Clarity	Clear	Clear

Microbiological Performance	Initial Inoculum	Test Result
-----------------------------	------------------	-------------

**Anaerobic incubation at 37°C for 48 hours**

<i>Peptostreptococcus anaerobius</i> ATCC®27337	48	Turbid growth
<i>Bacteroides thetaiotaomicron</i> ATCC®29148	1E+04-1E+06	Turbid growth
<i>Prevotella loescheii</i> ATCC®15930	1E+04-1E+06	Turbid growth
<i>Clostridium tetani</i> ATCC®9441	1E+04-1E+06	Turbid growth
<i>Clostridium sphenoides</i> ATCC®19403	1E+04-1E+06	Turbid growth
<i>Clostridium perfringens</i> ATCC®13124	1E+04-1E+06	Turbid growth
<i>Fusobacterium nucleatum</i> ATCC®10953	1E+04-1E+06	Turbid growth

Control Media: Tryptone Soya Agar or Columbia Blood Agar Base enriched with 5% v/v horse blood, where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu). For other positive test strains, a satisfactory result is represented by visible growth in accordance with the specification from an inoculum of 1E+04 to 1E+06 cfu. Refer to product specification for full details.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2023.09.29

.....  
Carissa Courtney  
Director Quality, EMEA



Tested by the Quality Control Laboratory  
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



## CERTIFICATE OF ANALYSIS

<b>PRODUCT</b>	<b>CM0957B</b>
	<b>ANAEROBE BASAL BROTH      500g</b>
<b>LOT NUMBER</b>	<b>3725312</b>

---

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**Tested by the Quality Control Laboratory**  
**OXOID LIMITED**

**Wade Road, Basingstoke, Hampshire RG24 8PW. England**  
**oxoid@thermofisher.com   www.oxoid.com**

FDA Reg No. 8010096