

CERTIFICATE OF ANALYSIS

PRODUCT CM1032B
HEART INFUSION BROTH 500g

LOT NUMBER 6196279

EXPIRY DATE 2029.11.30

DATE OF MANUFACTURE 2024.11.14

Delivery/Customer information

Date Printed

2025.01.22

Delivery No.

Customer

Customer Order Number

Physical Characteristics

Results

Specification

Appearance

Straw powder

Straw powder

Colour on reconstitution

Straw 3-4

Straw 3-4

pH (25°C)

7.3

7.2 - 7.6

Clarity

Clear

Clear

Microbiological Performance

Initial
Inoculum

Test Result

Aerobic incubation at 37 ± 2°C for 18 hours

Streptococcus pyogenes ATCC®19615

18

Turbid growth

Streptococcus pneumoniae ATCC®6303

35

Turbid growth

Streptococcus pneumoniae ATCC®6305

19

Turbid growth

Enterococcus faecalis ATCC®19433

31

Turbid growth

Pseudomonas aeruginosa ATCC®27853

39

Turbid growth

Anaerobic incubation at 37 ± 2°C for 18 hours

Streptococcus pneumoniae ATCC®6305

19

Turbid growth

Aerobic incubation at 37 ± 2°C for 48 hours

Candida albicans ATCC®10231

40

Turbid growth

Addition of 0.1% w/v bacteriological agar

Aerobic incubation at 37 ± 2°C for 48-72 hours

Bacteroides fragilis ATCC®25285

48

Turbid growth

Enriched with X & V factor discs

Aerobic incubation at 37 ± 2°C for 24 hours

Haemophilus influenzae ATCC®33391

38

Turbid growth

Enriched with 10% v/v horse blood

Aerobic incubation at 37 ± 2°C for 48 hours and subculture onto chocolate agar

Neisseria meningitidis ATCC®13077

14

Grey/brown cols

Neisseria gonorrhoeae NCTC11148

72

Grey/brown cols

Neisseria gonorrhoeae ATCC®19424

79

Grey/brown cols

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT CM1032B
HEART INFUSION BROTH 500g
LOT NUMBER 6196279

Tube coagulase test

Aerobic incubation at 37 ± 2°C for 4 hours

<i>Staphylococcus aureus</i> ATCC®9144	1E+04-1E+06	Positive
<i>Staphylococcus epidermidis</i> ATCC®14990	1E+04-1E+06	Negative

Tested in accordance with current CLSI M22 A

Aerobic incubation at 33-37°C for 18 hours

<i>Staphylococcus aureus</i> ATCC®25923	57	Turbid growth
<i>Escherichia coli</i> ATCC®25922	36	Turbid growth

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood, Columbia Blood Agar Base enriched with 5% v/v chocolate horse blood or Sabouraud Dextrose Agar, where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu).

For *Neisseria* species, a satisfactory result is represented by a positive diagnostic reaction, on subculture, from an inoculum of 10-100 cfu.

For Tube coagulase test, a satisfactory result is represented by the appropriate coagulase reaction from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.01.20

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.