

CERTIFICATE OF ANALYSIS

PRODUCT CM1032B
HEART INFUSION BROTH 500g

LOT NUMBER 6230042

EXPIRY DATE 2030.02.28

DATE OF MANUFACTURE 2025.02.05

Delivery/Customer information

Date Printed

2025.03.15

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 3-4	Straw 3-4
pH (25°C)	7.2	7.2 - 7.6
Clarity	Clear	Clear

Microbiological Performance	Initial Inoculum	Test Result
-----------------------------	------------------	-------------

Aerobic incubation at 37 ± 2°C for 18 hours

<i>Streptococcus pyogenes</i> ATCC®19615	29	Turbid growth
<i>Streptococcus pneumoniae</i> ATCC®6303	30	Turbid growth
<i>Streptococcus pneumoniae</i> ATCC®6305	38	Turbid growth
<i>Enterococcus faecalis</i> ATCC®19433	34	Turbid growth
<i>Pseudomonas aeruginosa</i> ATCC®27853	56	Turbid growth

Anaerobic incubation at 37 ± 2°C for 18 hours

<i>Streptococcus pneumoniae</i> ATCC®6305	38	Turbid growth
---	----	---------------

Aerobic incubation at 37 ± 2°C for 48 hours

<i>Candida albicans</i> ATCC®10231	55	Turbid growth
------------------------------------	----	---------------

Addition of 0.1% w/v bacteriological agar

Aerobic incubation at 37 ± 2°C for 48-72 hours

<i>Bacteroides fragilis</i> ATCC®25285	85	Turbid growth
--	----	---------------

Enriched with X & V factor discs

Aerobic incubation at 37 ± 2°C for 24 hours

<i>Haemophilus influenzae</i> ATCC®33391	49	Turbid growth
--	----	---------------

Enriched with 10% v/v horse blood

Aerobic incubation at 37 ± 2°C for 48 hours and subculture onto chocolate agar

<i>Neisseria meningitidis</i> ATCC®13077	23	Grey/brown cols
<i>Neisseria gonorrhoeae</i> NCTC11148	97	Grey/brown cols
<i>Neisseria gonorrhoeae</i> ATCC®19424	65	Grey/brown cols

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com

CERTIFICATE OF ANALYSIS

PRODUCT CM1032B
HEART INFUSION BROTH 500g
LOT NUMBER 6230042

Tube coagulase test

Aerobic incubation at 37 ± 2°C for 4 hours

<i>Staphylococcus aureus</i> ATCC®9144	1E+04-1E+06	Positive
<i>Staphylococcus epidermidis</i> ATCC®14990	1E+04-1E+06	Negative

Tested in accordance with current CLSI M22 A

Aerobic incubation at 33-37°C for 18 hours

<i>Staphylococcus aureus</i> ATCC®25923	77	Turbid growth
<i>Escherichia coli</i> ATCC®25922	40	Turbid growth

Control Media: Tryptone Soya Agar, Columbia Blood Agar Base enriched with 5% v/v horse blood, Columbia Blood Agar Base enriched with 5% v/v chocolate horse blood or Sabouraud Dextrose Agar, where appropriate

A satisfactory result is represented by visible growth from an inoculum of 10-100 colony-forming units (cfu).

For *Neisseria* species, a satisfactory result is represented by a positive diagnostic reaction, on subculture, from an inoculum of 10-100 cfu.

For Tube coagulase test, a satisfactory result is represented by the appropriate coagulase reaction from an inoculum of 1E+04 to 1E+06 cfu.

Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.03.10

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.