

CERTIFICATE OF ANALYSIS

Tested in accordan	ce with current CLSI metho	dology Zone Size	Limits
Assay (mcg/disc)		26.1	
Microbiological Pe	rformance		
DATE OF MANUFACTURE	2022.09.21		
EXPIRY DATE	2025.09.30		Customer Order Number
LOT NUMBER	3552268		Customer
PRODUCT	CT0051B RL25 SULPHAMETHOXAZOLE		Date Printed 2022.09.29 Delivery No.
			Delivery/Customer information

	(mm)	(mm)	
Staphylococcus aureus ATCC®25923	17	15 - 21	
Enterococcus faecalis ATCC®29212	0	0 - 0	
Haemophilus influenzae ATCC®49247	0	0 - 0	

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct, all results reported inthis certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release. Lot Accepted. 2022.09.29

Glowmen

.....

Carissa Courtney Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection. NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

