

## CERTIFICATE OF ANALYSIS

PRODUCT CT0193B

**CFP30 CEFOPERAZONE** 

**LOT NUMBER** 3768927

**EXPIRY DATE** 2025.12.31

**DATE OF** 

**MANUFACTURE** 2023.12.08

Delivery/Customer information

Date Printed 2023.12.13 Delivery No.

Customer

Customer Order Number

## Microbiological Performance

Assay (mcg/disc) 32.6

## Tested in accordance with current CLSI methodology

	Zone Size	Limits
	(mm)	(mm)
Staphylococcus aureus ATCC®25923	28	22 - 28
Escherichia coli ATCC®25922	26	24 - 30
Pseudomonas aeruginosa ATCC®27853	22	19 - 25

Quality control ranges are internally derived based on historical data.

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.

Lot Accepted. 2023.12.13

Glarmen

Carissa Courtney

Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

