



# CERTIFICATE OF ANALYSIS

**PRODUCT** CT0263B  
CFS30 CEFSULODIN

**LOT NUMBER** 3397030

**EXPIRY DATE** 2022.12.31

**DATE OF MANUFACTURE** 2021.11.18

Delivery/Customer information	
Date Printed	2021.11.26
Delivery No.	
Customer	
Customer Order Number	

## Microbiological Performance

Assay (mcg/disc) 28.6

### Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	28	22 - 28
<i>Escherichia coli</i> ATCC®25922	10	6 - 13
<i>Pseudomonas aeruginosa</i> ATCC®27853	28	23 - 28

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed. Control ranges are internally derived based on historical data.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2021.11.26

.....  
Carissa Courtney  
Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.  
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory  
OXOID LIMITED  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com