

**CERTIFICATE OF ANALYSIS**

PRODUCT CT0424B
NET10 NETILMICIN

LOT NUMBER 1345472

EXPIRY DATE 2016.05.31

**DATE OF
MANUFACTURE** 2013.05.10

Delivery/Customer information

Date Printed

2022.11.23

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 11.9

Tested in accordance with current BSAC methodology


	Zone Size (mm)	Limits (mm)
<i>Escherichia coli</i> ATCC®25922	24	22 - 26
<i>Escherichia coli</i> ATCC®10536	24	22 - 27
<i>Pseudomonas aeruginosa</i> ATCC®27853	24	20 - 24
<i>Pseudomonas aeruginosa</i> ATCC®25668	20	17 - 20
<i>Staphylococcus aureus</i> ATCC®25923	25	22 - 28

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2013.05.14


.....
Carol MacBride
Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**Tested by the Product Performance Laboratory
OXOID LIMITED**

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096