ThermoFisher
SCIENTIFIC

OXOID

CERTIFICATE OF ANALYSIS

PRODUCT	**CT0434B** **NOR10 NORFLOXACIN**
LOT NUMBER	6257180
EXPIRY DATE	2028.04.13
DATE OF MANUFACTURE	2025.04.14

Delivery/Customer information
Date Printed
2025.05.02
Delivery No.

Customer
Customer Order Number

Microbiological Performance

Assay (mcg/disc)   12.3

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
Staphylococcus aureus ATCC®25923	22	17 - 28
Escherichia coli ATCC®25922	29	28 - 35
Pseudomonas aeruginosa ATCC®27853	27	22 - 29
Streptococcus pneumoniae ATCC®49619	20	15 - 21

Tested in accordance with current EUCAST methodology

Staphylococcus aureus ATCC®29213	20	18 - 24
Escherichia coli ATCC®25922	29	28 - 35
Streptococcus pneumoniae ATCC®49619	22	18 - 24
Enterococcus faecalis ATCC®29212	19	16 - 22

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.02

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.