

**CERTIFICATE OF ANALYSIS**

PRODUCT CT0477B
ZOX30 CEFTIZOXIME

LOT NUMBER 1925932

EXPIRY DATE 2019.09.30

**DATE OF
MANUFACTURE** 2016.09.08

Delivery/Customer information

Date Printed

2018.07.11

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 30.75

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	27	27 - 35
<i>Escherichia coli</i> ATCC®25922	30	30 - 36
<i>Pseudomonas aeruginosa</i> ATCC®27853	16	12 - 17
<i>Neisseria gonorrhoeae</i> ATCC®49226	46	42 - 51
<i>Haemophilus influenzae</i> ATCC®49247	33	29 - 39
<i>Streptococcus pneumoniae</i> ATCC®49619	33	28 - 34

Tested in accordance with current BSAC methodology

<i>Escherichia coli</i> ATCC®10536	44	44 - 49
<i>Escherichia coli</i> ATCC®25922	43	40 - 44

Tested in accordance with current BSAC methodology

<i>Streptococcus pneumoniae</i> ATCC®49619	36	36 - 44
--------------------------------------------	----	---------

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.



Tested by the Quality Control Laboratory
OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096



CERTIFICATE OF ANALYSIS

PRODUCT CT0477B
LOT NUMBER ZOX30 CEFTIZOXIME
1925932

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke.

The results reported were obtained at the time of release.

Lot Accepted. 2016.09.12

Shannon Bearpark
Site Quality Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



Tested by the Quality Control Laboratory
OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com

FDA Reg No. 8010096