

CERTIFICATE OF ANALYSIS

PRODUCT CT0477B

ZOX30 CEFTIZOXIME

LOT NUMBER 3259905

EXPIRY DATE 2024.03.10

DATE OF

MANUFACTURE 2021.03.11

Delivery/Customer information

Date Printed 2021.03.18 Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 31.73

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits	
		(mm)	
Staphylococcus aureus ATCC®25923	27	27 - 35	
Escherichia coli ATCC®25922	31	30 - 36	
Pseudomonas aeruginosa ATCC®27853	12	12 - 17	
Neisseria gonorrhoeae ATCC®49226	42	42 - 51	
Haemophilus influenzae ATCC®49247	35	29 - 39	
Streptococcus pneumoniae ATCC®49619	33	28 - 34	

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Quality Control Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.

Lot Accepted. 2021.03.18

CHarmen

.....

Carissa Courtney

Director Quality, EMEA

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

