

CERTIFICATE OF ANALYSIS

PRODUCT CT0522B

MUP5 MUPIROCIN

LOT NUMBER 1785737

EXPIRY DATE 2018.12.31

DATE OF

MANUFACTURE 2015.12.08

Delivery/Customer information

Date Printed 2018.07.11 Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 4.74

Tested in accordance with current BSAC methodology

	Zone Size (mm)	Limits
		(mm)
Staphylococcus aureus ATCC®9144	31	26 - 35
Staphylococcus aureus ATCC®25923	29	24 - 34

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. This certificate shall not be reproduced except in full, without written approval of the Product Performance Laboratory, Oxoid Limited, Basingstoke. The results reported were obtained at the time of release.

Lot Accepted. 2015.12.11

Carol MacBride

Product Performance Manager

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

